



#### CN0565 – Multichannel Bipolar Impedancemeter Evaluation

Dr. Ing. Antonio Dell'Osa and Msc. Gerardo Ames Lastra Laboratorio de Electrónica Aplicada y Biomedicina – UNTDF Argentina






















































The objective of this work is to evaluate the traceability, precision, and accuracy of the bipolar measurements taken with the CN0565 as a multichannel bipolar device. Measurements taken with the AD5933EBZ board on the same resistor values were used as the gold standard.





# Methodology


- The selected commercial values were: 470  $\Omega$ , 680  $\Omega$ , 1000  $\Omega$ , and 2200  $\Omega$ .



- First, each resistor in each array was measured using the AD5933EBZ board. Each resistor was measured three times.
- Second, the resistor array was placed on the multichannel output of the CN0565, and sequential measurements of each element in the array were taken. Each sequential measurement of each array element was repeated five times.







Measurements on 470  $\Omega$  resistor for (a) 5 kHz, (b) 10 kHz, (c) 50 kHz and (d) 75 kHz.














Measurements on 680  $\Omega$  resistor for (a) 5 kHz, (b) 10 kHz, (c) 50 kHz and (d) 75 kHz.














Measurements on 1 k $\Omega$  resistor for (a) 5 kHz, (b) 10 kHz, (c) 50 kHz and (d) 75 kHz.














Measurements on 2,2 k $\Omega$  resistor for (a) 5 kHz, (b) 10 kHz, (c) 50 kHz and (d) 75 kHz.














Traceability analysis: Average relative error per board: on the x-axis (resistance rating, measurement frequency)













Nominal value	Mean	SD	CV [%]
470	0.534	0.148	27.70
680	0.249	0.102	40.85
1000	0.032	0.090	285.45
2200	0.220	0.099	45.09

Nominal Error for each resistor between the mean and standard deviation (SD) of all measurements between AD5933EBZ measurements. CV: variation coefficient.

# Discussion

- Repeated measurements caused a communication failure after 25 readings, likely due to insufficient memory management in the AD5940, resolved only by resetting the ADICUP3029.

- CN0565 board's accuracy depends on its fixed 1  $k\Omega$  calibration resistor (Rcal).

- Other values exhibited significant errors due to suboptimal calibration and system limitations.



## Conclusion

- CN0565 board is limited for its fixed Rcal=1kohm and for its GUI based on Python.
- In this context, CN0565 is not a good option for multichannel bipolar measurements in "plug and play mode".
- We couldn't study the "matrix switch" noise because the problem is measurement stage.



# Acknowledgments

- Professor Marco Zennaro (STI Unit, ICTP)
- Italian Embassy in Buenos Aires
- Training and Research in Italian Laboratories (TRIL) at the International Centre for Theoretical Physics (ICTP)







## Thank you for your attention





scientific instruments







